

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3951

Name: RITTER2  
Sterilizer ID: V2313228  
09/18/2024 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.2 F 0.0  
2:00 119.1 F 0.2  
4:00 128.1 F 1.8  
6:00 155.3 F 4.9  
8:00 182.3 F 8.6  
10:00 204.1 F 12.8  
12:00 214.2 F 4.9  
14:00 224.6 F 8.5  
16:00 235.5 F 12.6  
18:00 234.5 F 9.2  
20:00 241.0 F 11.2  
22:00 244.4 F 12.9  
24:00 255.7 F 18.8  
26:00 265.8 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.9  
1:00 272.6 F 29.8  
2:00 271.7 F 28.8  
3:00 272.6 F 29.7  
4:00 271.9 F 29.0

Min 271.5 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 202.3 F 0.0  
10:00 220.8 F 0.0  
15:00 204.7 F 0.0  
20:00 230.1 F 0.0  
25:00 227.9 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 27:41  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:06:18  
CYCLE COMPLETE

09/18/2024 12:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_