

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005380  
10:41 19/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:03  
STERILIZING 2:55  
136.4 C 222kPa 2:55  
136.3 C 223kPa 3:25  
136.4 C 223kPa 3:55  
136.4 C 222kPa 4:25  
136.0 C 218kPa 4:55  
136.4 C 222kPa 5:25  
136.3 C 224kPa 5:55  
136.4 C 225kPa 6:25  
136.4 C 225kPa 6:25  
VENTING 6:25  
AIR DRYING 7:18  
CYCLE COMPLETE 17:05

-----  
Digital Signature #  
C9D4F9FAB8026D4D  
-----