

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3358

Name: RITTER 1  
Sterilizer ID: V2178198  
06/08/2023 12:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.5 F 0.0  
2:00 173.4 F 0.0  
4:00 189.5 F 2.5  
6:00 207.6 F 7.3  
8:00 223.1 F 12.4  
10:00 228.8 F 8.0  
12:00 241.5 F 13.5  
14:00 241.3 F 11.5  
16:00 239.6 F 10.3  
18:00 253.7 F 17.2  
20:00 265.6 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.8 F 29.1  
2:00 272.3 F 28.7  
3:00 271.5 F 28.1  
4:00 272.1 F 28.8  
  
Min 271.4 F 28.1  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
5:00 217.5 F 0.0  
10:00 229.7 F 0.0  
15:00 229.4 F 0.0  
20:00 229.2 F 0.0  
25:00 229.8 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 21:48  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:59:56  
CYCLE COMPLETE

06/08/2023 01:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_