
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001993
12:33 09/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:27
STERILIZING 4:32
136.0 C 223kPa 4:32
136.2 C 222kPa 5:02
136.4 C 222kPa 5:32
136.2 C 223kPa 6:02
136.3 C 223kPa 6:32
136.4 C 225kPa 7:02
136.3 C 223kPa 7:32
136.4 C 222kPa 8:02
136.4 C 222kPa 8:02
VENTING 8:02
AIR DRYING 8:55
CYCLE COMPLETE 38:56

Digital Signature #
3BD2960445B771AA
