
Midmark M11 - v1.0.5
Total Cycles: 800

Name: RITTER2
Sterilizer ID: V2313228
02/15/2022 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 177.6 F 2.3
4:00 203.0 F 7.4
6:00 221.9 F 13.2
8:00 232.5 F 10.2
10:00 234.4 F 8.2
12:00 239.9 F 10.0
14:00 250.1 F 15.7
16:00 263.2 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.1 F 29.1
2:00 271.7 F 29.4
3:00 271.9 F 28.9
4:00 271.8 F 29.5

Min 271.3 F 28.6
Max 272.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 222.2 F 0.0
10:00 233.3 F 0.0
15:00 216.9 F 0.0
20:00 214.6 F 0.0
25:00 225.0 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 17:59
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

02/15/2022 11:06 AM
ACCEPTED: X
REJECTED:
Operator: U02