

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1612

Name: RITTER 1  
Sterilizer ID: V2178198  
02/09/2022 02:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.5 F 0.0  
2:00 78.1 F 0.0  
4:00 103.8 F 0.1  
6:00 139.5 F 1.9  
8:00 154.4 F 4.6  
10:00 172.1 F 8.0  
12:00 194.9 F 11.8  
14:00 211.1 F 3.6  
16:00 222.5 F 7.3  
18:00 234.5 F 11.1  
20:00 232.3 F 6.8  
22:00 241.0 F 12.0  
24:00 238.3 F 9.7  
26:00 251.0 F 15.9  
28:00 261.9 F 22.2  
30:00 272.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.4 F 29.2  
2:00 271.2 F 28.5  
3:00 272.4 F 29.3  
4:00 271.7 F 28.6

Min 271.2 F 28.4  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 225.1 F 0.0  
10:00 239.6 F 0.0  
15:00 229.3 F 0.0  
20:00 226.8 F 0.0  
25:00 228.1 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 30:30  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

02/09/2022 03:37 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_