
Midmark M11 - v1.0.5
Total Cycles: 4368

Name: RITTER2
Sterilizer ID: V2313228
02/21/2025 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.4 F 0.2
4:00 78.4 F 1.2
6:00 120.6 F 3.3
8:00 145.9 F 6.5
10:00 170.1 F 10.6
12:00 196.1 F 4.4
14:00 209.5 F 5.3
16:00 222.2 F 8.8
18:00 232.8 F 13.0
20:00 233.8 F 9.3
22:00 239.0 F 10.2
24:00 245.7 F 13.4
26:00 257.6 F 20.0
28:00 268.6 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 273.0 F 29.7
2:00 271.7 F 28.9
3:00 272.9 F 29.8
4:00 272.0 F 28.9

Min 271.6 F 28.8
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 185.8 F 0.0
10:00 209.9 F 0.0
15:00 208.4 F 0.0
20:00 207.6 F 0.0
25:00 205.4 F 0.0
30:00 203.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:12
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:07:26
CYCLE COMPLETE

02/21/2025 10:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____