
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001782
11:56 22/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:58
STERILIZING 4:04
136.0 C 220kPa 4:04
136.4 C 224kPa 4:35
136.4 C 225kPa 5:05
136.4 C 222kPa 5:35
136.3 C 222kPa 6:05
136.3 C 224kPa 6:35
136.3 C 224kPa 7:05
136.3 C 224kPa 7:35
136.3 C 224kPa 7:35
VENTING 7:35
AIR DRYING 8:26
CYCLE COMPLETE 8:31

Digital Signature #
5DCB98F0EA2DFEC6
