

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1718

Name: RITTER 1  
Sterilizer ID: V2178198  
03/15/2022 11:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
2:00 187.6 F 2.7  
4:00 189.8 F 6.6  
6:00 206.0 F 11.6  
8:00 219.5 F 5.8  
10:00 233.0 F 11.3  
12:00 234.2 F 8.3  
14:00 241.2 F 10.5  
16:00 245.9 F 13.3  
18:00 258.5 F 20.1  
20:00 269.8 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.1  
2:00 271.6 F 28.4  
3:00 272.3 F 29.2  
4:00 272.2 F 28.8

Min 271.4 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 229.2 F 0.0  
10:00 234.8 F 0.0  
15:00 227.2 F 0.0  
20:00 229.7 F 0.0  
25:00 230.1 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 20:59  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:59:15  
CYCLE COMPLETE

03/15/2022 12:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_