
Midmark M11 - v1.0.5
Total Cycles: 1138

Name: RITTER2
Sterilizer ID: V2313228
05/30/2022 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.5 F 0.0
2:00 135.7 F 1.5
4:00 152.5 F 4.6
6:00 178.2 F 8.8
8:00 207.7 F 12.4
10:00 219.2 F 7.8
12:00 230.7 F 12.7
14:00 232.0 F 9.4
16:00 237.3 F 9.9
18:00 245.3 F 13.6
20:00 256.6 F 20.0
22:00 267.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.6 F 30.4
2:00 272.4 F 29.9
3:00 271.4 F 29.5
4:00 272.4 F 30.4

Min 271.4 F 29.5
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 205.5 F 0.0
10:00 227.9 F 0.0
15:00 250.8 F 0.0
20:00 232.5 F 0.0
25:00 222.9 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:29
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:37
CYCLE COMPLETE

05/30/2022 01:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____