
Midmark M11 - v1.0.0
Total Cycles: 4258

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2024 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 179.4 F 3.1
4:00 194.2 F 7.4
6:00 209.8 F 12.7
8:00 225.2 F 7.0
10:00 239.0 F 12.2
12:00 236.0 F 9.4
14:00 237.8 F 9.2
16:00 248.5 F 14.4
18:00 260.6 F 21.1
20:00 271.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.7 F 29.1
2:00 272.2 F 28.6
3:00 271.3 F 28.1
4:00 272.1 F 29.0

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 218.8 F 0.0
10:00 216.7 F 0.0
15:00 212.4 F 0.0
20:00 201.2 F 0.0
25:00 194.3 F 0.0
30:00 191.2 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

02/01/2024 12:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____