

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007951  
9:18 09/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:05  
STERILIZING 3:05  
136.4 C 222kPa 3:05  
136.4 C 224kPa 3:35  
136.4 C 222kPa 4:05  
136.4 C 220kPa 4:35  
136.0 C 221kPa 5:05  
136.4 C 223kPa 5:35  
136.4 C 219kPa 6:05  
136.4 C 222kPa 6:35  
136.4 C 222kPa 6:35  
VENTING 6:35  
AIR DRYING 7:27  
CYCLE COMPLETE 37:27

-----  
Digital Signature #  
DB4199CC18484190  
-----