
Midmark M11 - v1.0.0
Total Cycles: 3235

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2023 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 188.0 F 0.9
4:00 212.0 F 5.2
6:00 230.6 F 10.6
8:00 225.2 F 5.8
10:00 238.2 F 11.6
12:00 235.3 F 8.9
14:00 239.0 F 9.7
16:00 247.5 F 13.9
18:00 259.7 F 20.5
20:00 270.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 272.4 F 29.2
2:00 272.0 F 29.0
3:00 271.8 F 28.6
4:00 272.4 F 28.7

Min 270.9 F 27.9
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.0 F 0.0
5:00 220.9 F 0.0
10:00 229.9 F 0.0
15:00 236.6 F 0.0
20:00 237.3 F 0.0
25:00 224.4 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

05/15/2023 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____