
Midmark M11 - v1.0.0
Total Cycles: 1998

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2022 06:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.9 F 0.0
2:00 77.8 F 0.0
4:00 86.3 F 0.2
6:00 115.0 F 2.6
8:00 145.2 F 5.8
10:00 201.3 F 9.1
12:00 229.2 F 12.9
14:00 220.4 F 4.9
16:00 230.2 F 8.3
18:00 239.0 F 12.1
20:00 232.4 F 7.6
22:00 243.7 F 12.8
24:00 239.8 F 10.3
26:00 252.2 F 16.4
28:00 262.7 F 22.6
30:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 271.7 F 28.6
2:00 272.3 F 28.9
3:00 272.6 F 29.1
4:00 272.2 F 29.3

Min 271.1 F 28.6
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 239.3 F 0.0
10:00 205.0 F 0.0
15:00 171.5 F 0.0
20:00 175.6 F 0.0
25:00 174.4 F 0.0
30:00 172.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 31:02
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

06/07/2022 08:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____