
Midmark M11 - v1.0.5
Total Cycles: 621

Name: RITTER2
Sterilizer ID: V2313228
12/09/2021 12:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.9 F 0.0
2:00 171.4 F 3.7
4:00 186.2 F 8.3
6:00 209.0 F 10.9
8:00 228.0 F 8.3
10:00 238.5 F 11.8
12:00 240.2 F 11.5
14:00 238.9 F 10.2
16:00 253.1 F 17.1
18:00 264.3 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.5
2:00 271.9 F 28.7
3:00 272.1 F 29.5
4:00 272.8 F 29.4

Min 271.3 F 28.6
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 234.3 F 0.0
10:00 230.9 F 0.0
15:00 230.5 F 0.0
20:00 223.6 F 0.0
25:00 216.2 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

12/09/2021 01:57 PM
ACCEPTED: X
REJECTED:
Operator: U02