
Midmark M11 - v1.0.5
Total Cycles: 1977

Name: RITTER2
Sterilizer ID: V2313228
01/30/2023 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.4 F 0.0
2:00 179.1 F 3.3
4:00 182.8 F 7.5
6:00 204.8 F 12.8
8:00 219.9 F 7.2
10:00 233.9 F 12.5
12:00 236.0 F 10.0
14:00 236.4 F 8.7
16:00 249.3 F 15.5
18:00 261.3 F 22.3
20:00 272.3 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.5 F 29.6
2:00 271.5 F 29.1
3:00 272.6 F 29.7
4:00 271.7 F 29.1

Min 271.5 F 29.0
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 217.0 F 0.0
10:00 225.8 F 0.0
15:00 228.0 F 0.0
20:00 233.5 F 0.0
25:00 229.0 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

01/30/2023 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____