

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4489

Name: RITTER 1  
Sterilizer ID: V2178198  
04/24/2024 04:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.2 F 0.0  
2:00 172.0 F 3.3  
4:00 189.6 F 7.4  
6:00 206.8 F 12.4  
8:00 222.5 F 6.7  
10:00 233.9 F 11.6  
12:00 234.2 F 8.6  
14:00 240.8 F 10.6  
16:00 246.5 F 13.5  
18:00 258.8 F 20.1  
20:00 269.7 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.2  
1:00 272.1 F 29.0  
2:00 272.0 F 28.5  
3:00 271.5 F 28.6  
4:00 272.4 F 28.9  
  
Min 271.2 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 224.2 F 0.0  
10:00 225.6 F 0.0  
15:00 234.4 F 0.0  
20:00 230.6 F 0.0  
25:00 230.8 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 20:52  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:00:19  
CYCLE COMPLETE

04/24/2024 05:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_