
Midmark M11 - v1.0.5
Total Cycles: 851

Name: RITTER2
Sterilizer ID: V2313228
03/01/2022 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 159.6 F 3.0
4:00 180.6 F 6.9
6:00 201.5 F 12.0
8:00 222.6 F 6.6
10:00 232.5 F 12.0
12:00 235.4 F 9.4
14:00 237.6 F 9.5
16:00 248.1 F 14.8
18:00 260.9 F 21.9
20:00 272.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.6 F 29.4
2:00 271.8 F 29.7
3:00 271.8 F 29.2
4:00 271.8 F 29.8

Min 271.0 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 227.9 F 0.0
10:00 217.3 F 0.0
15:00 214.5 F 0.0
20:00 205.5 F 0.0
25:00 206.4 F 0.0
30:00 214.2 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

03/01/2022 12:59 PM
ACCEPTED: X
REJECTED:
Operator: U02