
Midmark M11 - v1.0.0
Total Cycles: 4410

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2024 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.4 F 0.0
2:00 131.7 F 1.4
4:00 139.5 F 4.1
6:00 166.9 F 6.8
8:00 189.9 F 10.7
10:00 211.6 F 4.4
12:00 214.7 F 6.6
14:00 226.3 F 10.4
16:00 232.0 F 8.4
18:00 237.1 F 11.0
20:00 237.4 F 8.5
22:00 248.2 F 14.5
24:00 259.0 F 20.5
26:00 268.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.4 F 28.9
2:00 272.3 F 29.2
3:00 271.7 F 28.7
4:00 271.0 F 28.5

Min 271.0 F 28.4
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 213.1 F 0.0
10:00 228.8 F 0.0
15:00 230.4 F 0.0
20:00 229.9 F 0.0
25:00 226.2 F 0.0
30:00 224.7 F 0.0

Drying Complete

FILLING: 0:58
HEATING: 27:08
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:05:58
CYCLE COMPLETE

03/27/2024 12:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____