
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007577
13:28 19/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:34
STERILIZING 3:35
136.4 C 224kPa 3:35
136.3 C 224kPa 4:05
136.2 C 218kPa 4:35
136.4 C 222kPa 5:05
136.4 C 221kPa 5:35
136.2 C 223kPa 6:05
136.4 C 223kPa 6:35
136.5 C 223kPa 7:05
136.5 C 223kPa 7:05
VENTING 7:05
AIR DRYING 7:56
CYCLE COMPLETE 8:09

Digital Signature #
8B9E80877F00D053
