
Midmark M11 - v1.0.5
Total Cycles: 4353

Name: RITTER2
Sterilizer ID: V2313228
02/12/2025 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.8 F 0.0
2:00 134.6 F 1.7
4:00 153.5 F 4.6
6:00 179.5 F 8.7
8:00 201.3 F 13.4
10:00 220.4 F 5.9
12:00 232.1 F 10.4
14:00 230.4 F 7.8
16:00 241.7 F 12.4
18:00 240.3 F 10.9
20:00 253.2 F 17.3
22:00 264.4 F 24.1
24:00 271.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.9 F 29.8
2:00 272.4 F 29.1
3:00 271.5 F 29.0
4:00 272.8 F 29.8

Min 271.4 F 28.7
Max 273.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 189.2 F 0.0
10:00 193.9 F 0.0
15:00 196.3 F 0.0
20:00 196.4 F 0.0
25:00 194.8 F 0.0
30:00 193.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 24:03
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:31
CYCLE COMPLETE

02/12/2025 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____