

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 006156  
14:53 17/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:32  
STERILIZING 3:33  
136.4 C 223kPa 3:33  
136.3 C 224kPa 4:03  
136.1 C 223kPa 4:33  
136.1 C 217kPa 5:03  
136.4 C 224kPa 5:33  
136.4 C 221kPa 6:03  
136.3 C 225kPa 6:33  
136.4 C 224kPa 7:03  
136.4 C 224kPa 7:03  
VENTING 7:03  
AIR DRYING 7:53  
CYCLE COMPLETE 9:40

-----  
Digital Signature #  
2838554B70DA063B  
-----