
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
3.1uS / 1.9ppm
CYCLE NUMBER 003266
10:06 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:17
PRESSURIZING 2:23
STERILIZING 3:19
136.0 C 220kPa 3:19
136.4 C 221kPa 3:49
136.1 C 222kPa 4:19
136.0 C 221kPa 4:49
135.8 C 219kPa 5:19
136.2 C 222kPa 5:49
136.0 C 220kPa 6:19
136.4 C 222kPa 6:49
136.4 C 222kPa 6:49
VENTING 6:49
AIR DRYING 7:38
CYCLE COMPLETE 8:14

Digital Signature #
6963A31EE788ECBB
