
Midmark M11 - v1.0.5
Total Cycles: 1019

Name: RITTER2
Sterilizer ID: V2313228
04/21/2022 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.8 F 0.0
2:00 171.9 F 0.9
4:00 200.6 F 5.8
6:00 219.7 F 11.4
8:00 227.5 F 7.6
10:00 241.4 F 13.5
12:00 241.9 F 11.9
14:00 241.5 F 11.3
16:00 255.3 F 18.4
18:00 267.2 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.2
2:00 271.5 F 28.8
3:00 272.7 F 29.4
4:00 271.7 F 28.6

Min 271.4 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 228.5 F 0.0
10:00 222.2 F 0.0
15:00 233.1 F 0.0
20:00 221.7 F 0.0
25:00 222.4 F 0.0
30:00 217.9 F 0.0

Drying Complete

FILLING: 1:42
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

04/21/2022 12:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____