

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005965  
9:34 13/07/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:22  
STERILIZING 3:18  
136.4 C 220kPa 3:18  
136.3 C 224kPa 3:48  
136.4 C 224kPa 4:18  
136.4 C 221kPa 4:48  
136.4 C 223kPa 5:18  
136.4 C 222kPa 5:48  
136.4 C 224kPa 6:18  
136.4 C 222kPa 6:48  
136.4 C 222kPa 6:48  
VENTING 6:48  
AIR DRYING 7:37  
CYCLE COMPLETE 23:50

-----  
Digital Signature #  
F18C1BF0F7583555  
-----