
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001356
10:30 23/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:54
STERILIZING 4:59
136.0 C 220kPa 4:59
136.4 C 223kPa 5:29
136.1 C 220kPa 5:59
136.0 C 222kPa 6:29
136.1 C 222kPa 6:59
136.4 C 223kPa 7:29
136.4 C 225kPa 7:59
136.3 C 223kPa 8:29
136.3 C 223kPa 8:29
VENTING 8:29
AIR DRYING 9:24
CYCLE COMPLETE 24:24

Digital Signature #
7C838D9CE63A3495
