
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001530
13:58 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:06
STERILIZING 4:10
136.5 C 226kPa 4:10
136.2 C 224kPa 4:40
136.1 C 223kPa 5:10
136.4 C 223kPa 5:40
136.1 C 223kPa 6:10
136.3 C 225kPa 6:40
136.4 C 225kPa 7:10
136.5 C 226kPa 7:40
136.5 C 226kPa 7:40
VENTING 7:40
AIR DRYING 8:36
CASSETTE REMOVED
CYCLE COMPLETE 24:24

Digital Signature #
6B3BE855725C5D08
