
Midmark M11 - v1.0.0
Total Cycles: 796

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2021 03:19 PM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.2 F 0.0
2:00 177.3 F 3.0
4:00 202.4 F 5.1
6:00 218.6 F 4.7
8:00 223.2 F 4.3
10:00 228.8 F 5.7
12:00 236.4 F 8.8
14:00 249.4 F 14.8

STERILIZING
mm:ss Degrees PSI
0:00 251.6 F 16.0
1:00 252.4 F 16.3
2:00 252.3 F 16.5
3:00 252.2 F 16.1
4:00 252.3 F 16.5
5:00 252.5 F 16.3
6:00 251.7 F 16.0
7:00 252.9 F 16.6
8:00 252.4 F 16.3
9:00 251.8 F 15.9
10:00 252.4 F 16.5
11:00 253.0 F 16.5
12:00 252.4 F 16.2
13:00 251.9 F 16.0
14:00 251.9 F 16.2
15:00 253.0 F 16.6
16:00 252.9 F 16.5
17:00 252.5 F 16.3
18:00 252.1 F 16.1
19:00 251.7 F 15.9
20:00 252.3 F 16.4
21:00 253.0 F 16.6
22:00 252.9 F 16.5
23:00 252.5 F 16.3
24:00 252.2 F 16.1
25:00 251.8 F 16.0
26:00 251.9 F 16.1
27:00 252.7 F 16.6
28:00 253.0 F 16.6
29:00 252.8 F 16.5
30:00 252.5 F 16.3

Min 251.6 F 15.9
Max 253.1 F 16.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 203.0 F 0.0
10:00 198.2 F 0.0
15:00 207.7 F 0.0
20:00 214.1 F 0.0
25:00 224.5 F 0.0
30:00 204.9 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 14:56
STERILIZING: 30:00

VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:20:16
CYCLE COMPLETE

04/21/2021 04:39 PM
ACCEPTED: X
REJECTED:
Operator: U01