

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001058  
8:28 30/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:52  
PRESSURIZING 3:54  
STERILIZING 5:01  
136.4 C 224kPa 5:01  
136.3 C 223kPa 5:31  
136.4 C 224kPa 6:01  
136.4 C 222kPa 6:31  
136.3 C 221kPa 7:01  
136.4 C 223kPa 7:31  
136.3 C 222kPa 8:01  
136.0 C 221kPa 8:31  
136.0 C 221kPa 8:31  
VENTING 8:31  
AIR DRYING 9:25  
CYCLE COMPLETE 9:26  
-----

Digital Signature #  
82914D081E653275  
-----