

-----  
Midmark M11 - v1.0.0  
Total Cycles: 956

Name: RITTER 1  
Sterilizer ID: V2178198  
07/05/2021 06:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.5 F 0.0  
2:00 73.7 F 0.0  
4:00 81.2 F 0.0  
6:00 100.9 F 0.1  
8:00 132.9 F 0.1  
10:00 166.9 F 0.2  
12:00 190.7 F 0.8  
14:00 210.4 F 3.4  
16:00 215.0 F 6.1  
18:00 223.7 F 9.1  
20:00 233.2 F 12.5  
22:00 232.8 F 8.1  
24:00 242.7 F 12.9  
26:00 240.2 F 10.5  
28:00 236.7 F 8.3  
30:00 249.7 F 14.8  
32:00 260.7 F 20.9  
34:00 270.7 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 27.9  
1:00 272.5 F 28.6  
2:00 271.4 F 27.8  
3:00 272.5 F 28.7  
4:00 271.8 F 28.0  
  
Min 271.3 F 27.8  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
5:00 198.5 F 0.0  
10:00 193.4 F 0.0  
15:00 190.7 F 0.0  
20:00 207.3 F 0.0  
25:00 225.6 F 0.0  
30:00 231.9 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 34:50  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:12:55  
CYCLE COMPLETE

07/05/2021 08:04 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01