
Midmark M11 - v1.0.0
Total Cycles: 2667

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2022 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.2 F 0.0
2:00 138.0 F 0.1
4:00 180.2 F 1.2
6:00 214.2 F 5.4
8:00 223.7 F 10.4
10:00 225.7 F 5.9
12:00 237.7 F 10.9
14:00 232.5 F 7.2
16:00 247.0 F 13.5
18:00 244.3 F 12.0
20:00 256.8 F 18.6
22:00 268.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.2 F 28.5
2:00 272.6 F 28.8
3:00 272.2 F 29.1
4:00 271.6 F 28.5

Min 271.6 F 28.4
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 222.2 F 0.0
10:00 232.0 F 0.0
15:00 226.8 F 0.0
20:00 226.8 F 0.0
25:00 228.9 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 24:00
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:14
CYCLE COMPLETE

12/13/2022 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____