
Midmark M11 - v1.0.0
Total Cycles: 2941

Name: RITTER 1
Sterilizer ID: V2178198
02/23/2023 04:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.5 F 0.0
2:00 169.0 F 4.0
4:00 188.3 F 8.5
6:00 207.5 F 13.4
8:00 221.1 F 7.6
10:00 234.5 F 12.9
12:00 236.3 F 10.0
14:00 236.0 F 8.3
16:00 249.2 F 15.0
18:00 261.7 F 22.1
20:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 272.6 F 29.2
4:00 271.8 F 28.6

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 231.3 F 0.0
10:00 226.1 F 0.0
15:00 225.7 F 0.0
20:00 227.6 F 0.0
25:00 229.4 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

02/23/2023 05:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____