

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2295

Name: RITTER2  
Sterilizer ID: V2313228  
04/24/2023 04:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
2:00 179.5 F 3.7  
4:00 189.2 F 8.4  
6:00 216.5 F 9.1  
8:00 228.5 F 9.5  
10:00 234.2 F 8.6  
12:00 243.3 F 12.6  
14:00 244.8 F 13.2  
16:00 258.0 F 20.3  
18:00 269.7 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.5 F 29.8  
2:00 271.2 F 29.0  
3:00 272.2 F 30.2  
4:00 272.9 F 30.1  
  
Min 271.0 F 29.0  
Max 273.3 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
5:00 221.9 F 0.0  
10:00 234.7 F 0.0  
15:00 231.0 F 0.0  
20:00 225.8 F 0.0  
25:00 230.2 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 18:57  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:04  
CYCLE COMPLETE

04/24/2023 05:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_