
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 003772
9:02 07/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:50
STERILIZING 3:56
136.0 C 220kPa 3:56
136.0 C 222kPa 4:27
136.3 C 223kPa 4:57
136.4 C 223kPa 5:27
136.4 C 223kPa 5:57
136.5 C 224kPa 6:27
136.3 C 224kPa 6:57
136.4 C 223kPa 7:27
136.4 C 223kPa 7:27
VENTING 7:27
AIR DRYING 8:20
CYCLE COMPLETE 8:40

Digital Signature #
4F0B816DC06F1696
