

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1642

Name: RITTER 1  
Sterilizer ID: V2178198  
02/16/2022 01:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.3 F 0.0  
2:00 130.0 F 1.2  
4:00 154.5 F 3.6  
6:00 182.1 F 7.5  
8:00 202.4 F 12.5  
10:00 217.1 F 5.6  
12:00 230.2 F 10.6  
14:00 231.9 F 7.1  
16:00 243.7 F 12.9  
18:00 242.2 F 11.4  
20:00 255.0 F 17.9  
22:00 266.4 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.8 F 29.1  
2:00 271.8 F 28.3  
3:00 272.2 F 28.9  
4:00 272.6 F 29.0

Min 271.5 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.6 F 0.0  
5:00 217.4 F 0.0  
10:00 225.2 F 0.0  
15:00 229.0 F 0.0  
20:00 226.2 F 0.0  
25:00 222.0 F 0.0  
30:00 221.6 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 23:35  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:41  
CYCLE COMPLETE

02/16/2022 02:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_