
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000499
10:46 07/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:17
STERILIZING 4:29
136.0 C 228kPa 4:29
136.4 C 232kPa 4:59
136.4 C 233kPa 5:29
136.2 C 230kPa 5:59
136.4 C 233kPa 6:29
136.4 C 231kPa 6:59
136.3 C 233kPa 7:29
136.4 C 233kPa 7:59
136.4 C 233kPa 7:59
VENTING 7:59
AIR DRYING 8:53
CASSETTE REMOVED
CYCLE COMPLETE 9:41

Digital Signature #
284D690A53B1AB13
