

-----  
Midmark M11 - v1.0.5  
Total Cycles: 128

Name: RITTER2  
Sterilizer ID: V2313228  
06/22/2021 04:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 96.6 F 0.0  
2:00 97.3 F 0.6  
4:00 118.7 F 2.2  
6:00 147.2 F 4.9  
8:00 173.9 F 8.4  
10:00 193.1 F 12.2  
12:00 209.6 F 4.8  
14:00 218.2 F 8.8  
16:00 229.7 F 13.1  
18:00 233.0 F 9.7  
20:00 238.8 F 10.1  
22:00 244.6 F 13.5  
24:00 255.6 F 19.4  
26:00 266.0 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.7  
1:00 272.3 F 30.4  
2:00 271.7 F 29.6  
3:00 272.2 F 30.4  
4:00 272.1 F 29.9

Min 271.3 F 29.5  
Max 272.8 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.7 F 0.0  
5:00 207.2 F 0.0  
10:00 215.9 F 0.0  
15:00 215.8 F 0.0  
20:00 214.2 F 0.0  
25:00 216.8 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 27:36  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:05:53  
CYCLE COMPLETE

06/22/2021 05:55 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_