
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000755
11:10 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:56
STERILIZING 4:03
136.0 C 230kPa 4:03
136.4 C 232kPa 4:33
136.4 C 233kPa 5:03
136.4 C 233kPa 5:33
136.4 C 233kPa 6:03
136.4 C 230kPa 6:33
136.3 C 233kPa 7:03
136.4 C 233kPa 7:33
136.4 C 233kPa 7:33
VENTING 7:33
AIR DRYING 8:29
CYCLE COMPLETE 8:31

Digital Signature #
BE38B5032593B0CB
