

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000972  
11:52 09/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:07  
PRESSURIZING 2:13  
STERILIZING 3:14  
136.4 C 233kPa 3:14  
136.3 C 231kPa 3:44  
136.3 C 231kPa 4:14  
136.4 C 231kPa 4:44  
136.3 C 231kPa 5:14  
136.4 C 233kPa 5:44  
136.0 C 231kPa 6:14  
135.9 C 230kPa 6:44  
135.9 C 230kPa 6:44  
VENTING 6:44  
AIR DRYING 7:36  
CYCLE COMPLETE 7:42  
-----

Digital Signature #  
E3853878832ECA30  
-----