
Midmark M11 - v1.0.5
Total Cycles: 1971

Name: RITTER2
Sterilizer ID: V2313228
01/25/2023 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.4 F 0.0
2:00 181.2 F 2.6
4:00 188.4 F 7.0
6:00 209.8 F 12.6
8:00 222.6 F 7.6
10:00 236.3 F 13.3
12:00 238.8 F 11.1
14:00 239.0 F 10.0
16:00 252.7 F 17.1
18:00 264.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.1 F 29.5
2:00 272.4 F 29.4
3:00 271.5 F 28.8
4:00 272.5 F 29.8

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 220.0 F 0.0
10:00 227.6 F 0.0
15:00 229.7 F 0.0
20:00 229.2 F 0.0
25:00 229.9 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

01/25/2023 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____