

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2625

Name: RITTER2  
Sterilizer ID: V2313228  
07/10/2023 09:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.1 F 0.0  
2:00 183.6 F 2.0  
4:00 181.6 F 5.9  
6:00 199.5 F 11.1  
8:00 223.8 F 6.1  
10:00 239.0 F 11.5  
12:00 234.5 F 8.8  
14:00 238.8 F 10.4  
16:00 246.7 F 13.8  
18:00 258.6 F 20.4  
20:00 269.4 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.8  
1:00 272.4 F 29.6  
2:00 271.5 F 28.7  
3:00 272.4 F 29.5  
4:00 271.5 F 28.7  
  
Min 271.2 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
5:00 228.1 F 0.0  
10:00 225.7 F 0.0  
15:00 218.4 F 0.0  
20:00 219.3 F 0.0  
25:00 219.6 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:46  
CYCLE COMPLETE

07/10/2023 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_