
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002532
12:03 21/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:51
STERILIZING 3:55
136.0 C 220kPa 3:55
136.2 C 224kPa 4:25
136.2 C 223kPa 4:55
136.4 C 225kPa 5:25
136.4 C 224kPa 5:55
136.2 C 223kPa 6:25
136.3 C 224kPa 6:55
136.4 C 223kPa 7:25
136.4 C 223kPa 7:25
VENTING 7:25
AIR DRYING 8:21
CYCLE COMPLETE 16:29

Digital Signature #
48F88F023097A89E
