

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002326  
15:55 21/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:57  
STERILIZING 2:54  
136.4 C 223kPa 2:54  
136.1 C 224kPa 3:24  
136.4 C 224kPa 3:54  
136.3 C 222kPa 4:24  
136.4 C 226kPa 4:54  
136.4 C 223kPa 5:24  
136.1 C 223kPa 5:54  
136.3 C 226kPa 6:24  
136.3 C 226kPa 6:24  
VENTING 6:24  
AIR DRYING 7:17  
CYCLE COMPLETE 37:17  
-----

Digital Signature #  
83077C8F2EC4B12E  
-----