
Midmark M11 - v1.0.0
Total Cycles: 5229

Name: RITTER 1
Sterilizer ID: V2178198
01/09/2025 09:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 76.5 F 0.0
4:00 79.4 F 0.0
6:00 104.2 F 0.1
8:00 140.0 F 0.2
10:00 171.4 F 0.4
12:00 197.0 F 0.8
14:00 215.6 F 4.4
16:00 223.8 F 7.5
18:00 232.5 F 11.2
20:00 230.3 F 6.7
22:00 243.5 F 12.6
24:00 240.6 F 10.7
26:00 237.6 F 9.2
28:00 252.5 F 16.3
30:00 264.3 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 271.8 F 27.9
2:00 272.1 F 28.1
3:00 272.4 F 28.4
4:00 272.7 F 28.6

Min 271.6 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 187.5 F 0.0
10:00 185.8 F 0.0
15:00 185.3 F 0.0
20:00 184.1 F 0.0
25:00 184.3 F 0.0
30:00 184.0 F 0.0
Drying Complete

FILLING: 0:42
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:60
CYCLE COMPLETE

01/09/2025 10:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____