
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002250
11:18 13/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:36
STERILIZING 3:29
136.0 C 220kPa 3:29
136.0 C 222kPa 3:59
136.3 C 224kPa 4:29
136.3 C 224kPa 4:59
136.2 C 223kPa 5:29
136.4 C 225kPa 5:59
136.3 C 222kPa 6:29
136.2 C 224kPa 6:59
136.2 C 224kPa 6:59
VENTING 6:59
AIR DRYING 7:49
CYCLE COMPLETE 28:33

Digital Signature #
653D9CCB647647D6
