
Midmark M11 - v1.0.0
Total Cycles: 2033

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2022 07:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.8 F 0.0
2:00 75.5 F 0.0
4:00 83.5 F 0.0
6:00 112.2 F 1.3
8:00 143.6 F 2.8
10:00 172.8 F 5.2
12:00 193.8 F 8.3
14:00 208.2 F 11.7
16:00 218.7 F 3.4
18:00 226.0 F 7.7
20:00 235.4 F 11.7
22:00 233.5 F 7.8
24:00 245.3 F 13.4
26:00 242.9 F 11.5
28:00 254.9 F 17.7
30:00 265.4 F 24.0
32:00 271.1 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.7 F 29.0
2:00 272.0 F 28.8
3:00 272.5 F 28.9
4:00 272.0 F 28.9

Min 271.7 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 209.5 F 0.0
10:00 199.7 F 0.0
15:00 193.0 F 0.0
20:00 203.6 F 0.0
25:00 218.4 F 0.0
30:00 217.9 F 0.0

Drying Complete

FILLING: 0:44
HEATING: 32:43
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:36
CYCLE COMPLETE

06/20/2022 08:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____