
Midmark M11 - v1.0.0
Total Cycles: 962

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2021 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.5 F 0.0
2:00 207.7 F 4.0
4:00 211.7 F 9.0
6:00 229.9 F 8.5
8:00 233.8 F 9.8
10:00 239.3 F 8.7
12:00 244.9 F 13.3
14:00 244.0 F 12.2
16:00 256.7 F 18.9
18:00 267.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.7
2:00 272.0 F 28.8
3:00 272.1 F 28.5
4:00 272.2 F 29.0

Min 271.5 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 217.4 F 0.0
10:00 226.6 F 0.0
15:00 228.6 F 0.0
20:00 229.9 F 0.0
25:00 228.5 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

07/05/2021 01:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____