

-----  
Midmark M11 - v1.0.5  
Total Cycles: 26

Name: RITTER2  
Sterilizer ID: V2313228  
04/27/2021 12:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 125.2 F 0.0  
2:00 129.8 F 1.5  
4:00 155.0 F 4.2  
6:00 175.3 F 8.0  
8:00 195.9 F 12.4  
10:00 210.4 F 5.7  
12:00 223.4 F 10.2  
14:00 230.8 F 8.7  
16:00 239.8 F 12.4  
18:00 240.8 F 11.3  
20:00 252.6 F 17.4  
22:00 263.3 F 23.7  
24:00 272.7 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.4 F 29.7  
2:00 271.7 F 29.4  
3:00 272.4 F 29.7  
4:00 271.6 F 29.3

Min 271.5 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 208.3 F 0.0  
10:00 207.1 F 0.0  
15:00 204.4 F 0.0  
20:00 202.0 F 0.0  
25:00 199.3 F 0.0  
30:00 196.1 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 24:16  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:02:53  
CYCLE COMPLETE

04/27/2021 01:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_