
Midmark M11 - v1.0.0
Total Cycles: 4560

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2024 07:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 73.1 F 0.0
4:00 81.9 F 0.2
6:00 116.2 F 1.9
8:00 147.4 F 4.7
10:00 176.0 F 7.6
12:00 213.9 F 11.5
14:00 214.2 F 3.4
16:00 222.8 F 6.8
18:00 228.1 F 10.6
20:00 231.7 F 8.1
22:00 238.9 F 10.5
24:00 235.4 F 7.9
26:00 247.4 F 13.8
28:00 258.5 F 19.9
30:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.7 F 28.8
2:00 271.4 F 28.2
3:00 272.7 F 28.9
4:00 271.6 F 28.1

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 234.8 F 0.0
10:00 198.9 F 0.0
15:00 184.9 F 0.0
20:00 193.7 F 0.0
25:00 196.9 F 0.0
30:00 210.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:11
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

05/16/2024 08:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____