

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.8uS / 1.1ppm  
CYCLE NUMBER 001874  
13:31 19/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:52  
STERILIZING 4:01  
136.3 C 222kPa 4:01  
136.2 C 221kPa 4:31  
136.1 C 223kPa 5:01  
136.3 C 224kPa 5:31  
136.4 C 226kPa 6:01  
136.4 C 223kPa 6:31  
136.4 C 224kPa 7:01  
136.3 C 224kPa 7:31  
136.3 C 224kPa 7:31  
VENTING 7:31  
AIR DRYING 8:19  
CYCLE COMPLETE 11:02  
-----

Digital Signature #  
78D2E973B0573882  
-----